

# Result Report r0000002

OptiFlash 000166

V 1.00

23 Mar 2015 09:27

Date: 24 Feb 2015 10:40 v1.00

Operator: PAC user

Sample ID: CRM 1045

Sample des.:

Note:

Standard: ASTM D93A

Product: Factory ASTM D93A °C

Expected flash point: 217.0°C

Uncorrected flash point: 214.0°C

Ambient pressure: 98.2kPa

Corrected flash point: 215.0°C

Ending: Stop on flash point

Alarm:

Status: Result ok

Min.: °C Max.: °C

Instrument name: 104004

Ignition type: Electric

Pretest: Off

Warnings:

Max sample temp: 214.3°C

Sample probe SN: 15020018

Correlation:

Bitumen: Off

## Heating rate graph

