

CERTIFICATE OF MEASUREMENT

Instructions for Use

Suitable for determination of flash/no flash - rapid equilibrium closed cup methods ASTM D3278; D3828; IP303 (obs); IP 523; IP524; ISO3679; ISO3680; BS3900 Part A13.

Suitable for determination of flash point - small scale closed cup ramp methods ASTM D7236; IP 534.

Certification Procedure

This certified reference material is produced in accordance with ISO Guides 34 and 35.

Certified Value

Certified Flash/No Flash Value = 74.5°C (ASTM D3278; D3828; IP 523; IP 524; ISO3679; ISO3680)

Certified Ramp Flash Point Value = 74.2°C (ASTM D7236; IP 534)

The certified value has an uncertainty of 0.5°C based on the 95% confidence limit for the mean.

Safety Precautions and Storage Requirements

It is the responsibility of the user of this verification material to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use. Supplied packed in accordance with IATA regulations for international transportation. The material should be stored in a cool place away from direct sunlight.

Batch No: 1615061c
.....

Approved Signatory:



Andrew Long