



est. 1878

# Certificate of Calibration

Tamson Instruments by states following:

Serial No.: 10CD07  
 Description: TT3 - Handheld  
 Part No.: 10T6094  
 Range: -40..+140 [°C]  
 Digits: 3  
 Date Batch: 1-4-2023  
 Immersion Depth: 120 [mm]

Reference Value		Deviation
[°C]	[°C]	[°C]
-20,011	-20,019	-0,008
0,023	0,019	-0,004
50,001	49,993	-0,008
100,002	99,996	-0,006
130,007	129,999	-0,008

Requested testpoints: N.A.

test points	9	#
Overall test range	-40 .. +140	[°C]
max deviation found	0,000	[°C]
min deviation found	-0,008	[°C]
delta deviation	0,008	[°C]
average deviation ±	0,004	[°C]
stDev	0,003	[°C]
Immersion depth	120	[mm]

At the time of manufacture, this certifies that the above product was calibrated in accordance with applicable Tamson procedures which are in compliance with relevant clause of ISO 9001 - 2015.

At the time of shipment, this instrument meets its published operating specifications. The environment in which this instrument was calibrated is kept within the operating specifications of the instrument and standards.

Supporting documentation relative to traceability has file number TD-13115126. The documentation is traceable to national standards.

The calibration period of this instrument begins on the date of shipment. Calibration interval is 12 months.

Performed by:



ISO 9001 : 2015  
NL/PRO 238239125

## Tamson Instruments by

E-mail: [info@tamson.com](mailto:info@tamson.com)  
Website: [www.tamson.com](http://www.tamson.com)