

Certificate of Verification

Foam Gas Diffuser

Diffuser Type: Cylindrical, Stainless Steel

Serial #: 8-34

Critical Measurements

Maximum Pore Diameter: 27.7 micrometers
422 mm of water

Acceptance Criteria:
(Not greater than 80 micrometers)

Permeability: 4490 mL air/min @ 250 mm water

Acceptance Criteria:
(3000 to 6000 mL of air/minute)

Acceptance Date

08-11-20

• *Re-verification should be performed every 3 months or ~100 tests, depending upon usage rate.*

• *Verification testing can be performed at Tannas Co.*

• *Call at 989-496-2309*

Calibration Equipment:

1000 mL U-tube Manometer
Digital Precision Flowmeter, Humonics Model 730



Tannas Co.
Quality Control

Marc [Signature]